



Australian Government

MSL933008 Perform calibration checks on equipment and assist with its maintenance

Release: 1

MSL933008 Perform calibration checks on equipment and assist with its maintenance

Modification History

Release	Comments
Release 1	<p>This version was released in <i>MSL Laboratory Operations Training Package Release 2.0</i>.</p> <p>Supersedes and equivalent to MSL933004 Perform calibration checks on equipment and assist with its maintenance. Range of conditions removed. Assessment requirements amended.</p>

Application

This unit of competency describes the skills and knowledge to perform set-up, pre-use and in-house calibration/validation checks on laboratory equipment and assist with laboratory equipment maintenance.

This unit of competency applies to laboratory assistants working in all industry sectors.

No licensing or certification requirements exist at the time of publication. However, regulations and/or external accreditation requirements for laboratory operations exist, so local requirements should be checked. Relevant legislation, industry standards and codes of practice within Australia must also be applied.

Pre-requisite Unit

Nil

Competency Field

Calibration

Elements and Performance Criteria

Elements describe the essential outcomes.

Performance criteria describe the performance needed to demonstrate achievement of the element.

1	Perform set-up and pre-use checks of laboratory	1.1	Perform laboratory equipment set-up and pre-use checks in accordance with workplace procedures
		1.2	Perform safety checks in accordance with relevant workplace and instrumental procedures

Elements describe the essential outcomes.

Performance criteria describe the performance needed to demonstrate achievement of the element.

- | | |
|--|--|
| equipment | 1.3 Identify faulty or unsafe components and equipment and report to appropriate personnel |
| | 1.4 Complete equipment log books/records to meet workplace requirements |
| 2 Perform calibration checks | 2.1 Start up equipment according to operating procedures |
| | 2.2 Use specified standards for calibration check |
| | 2.3 Check equipment in accordance with calibration procedures and schedules |
| | 2.4 Record all calibration data accurately and legibly |
| | 2.5 Compare data with specifications and/or previous records to identify non-compliant equipment |
| | 2.6 Quarantine out-of-calibration equipment |
| 3 Assist with equipment maintenance | 3.1 Ensure all equipment work areas are clean during and after equipment use |
| | 3.2 Perform basic maintenance in accordance with workplace procedures |
| | 3.3 Clean and store equipment according to workplace and/or manufacturer's specifications/procedures |
| | 3.4 Identify and replace, repair or dispose of damaged/worn equipment |
| 4 Maintain records | 4.1 Record maintenance details |
| | 4.2 Report information on unsafe or faulty equipment according to workplace procedures |

Foundation Skills

This section describes those language, literacy, numeracy and employment skills that are essential to performance.

Foundation skills essential to performance are explicit in the performance criteria of this unit of competency.

Unit Mapping Information

Equivalent to MSL933004 Perform calibration checks on equipment and assist with its maintenance, Release 1.

Links

MSL Laboratory Operations Companion Volume Implementation Guide is available from VETNet -

<https://vetnet.gov.au/Pages/TrainingDocs.aspx?q=5c63a03b-4a6b-4ae5-9560-1e3c5f462baa>